Arnd Meyer
Jul 06, 2000

Online Status – Integration

- Last Friday – most successful so far (as always...)
  - New L3 executable runs stable (Calorimetry, JetClu, CdfMet, JetFilterModule), no rejection
  - CMU vs. CCAL roughly timed in – no correlations yet
  - QIE calibration writes (again) directly to the database
  - After sorting out various problems, one very nice run 93549 with about 50 crates, 220k events on disk @ 70 Hz –
    - CCAL02–05, CCAL10–13, CMP, CMX, CMU (HV on), L1CL, L2CL (not 00), L1GL, Scalers, WCAL (not 05, 07), PCAL 00,01,05–10, + TSI/EVB/L3/errorHandler/.../YMon/EVD/...
  - Excluded due to various problems
    - COT crates (VRB link), IMU00 (bad data), XFT Linker (configuration problems), WCAL 07(bad data), CLC (used), ... Need to sort these out
- Other successes:
  - Control COT calibration from the (real) pulser crate
  - Ran (part of) TrigSim online (CalTrigger) online
- Working on, among other things ...
  - Understand run failures – bad data from crates etc.
  - Muon/COT/CLC calibration (+ others)
  - Run with full XFT
  - Establish muon trigger, coincidence with calorimeter trigger
  - Error reporting, handling, and recovery